

TEST REPORT

2020TM0361

DATE OF RECEPTION

19/03/2020

APPLICANT

DATE TESTS

Starting: 19/03/2020

Ending: 28/03/2020

IDENTIFICATION AND DESCRIPTION OF SAMPLES

REFERENCES

MASK REF: PTXMASKV2_HOSP

TESTS CARRIED OUT

- CARRIED OUT ON THE FOLLOWING REFERENCE:.
- IN VITRO DETERMINATION OF BACTERIAL FILTRATION EFFICIENCY (BFE)*.
- DETERMINATION OF BREATHABILITY (DIFFERENTIAL PRESSURE)*.
- DETERMINATION OF PRESSURE OF SPLASH RESISTANCE*.
- DETERMINATION OF A POPULATION OF MICROORGANISMS ON PRODUCTS.
- TEST FOR CYTOTOXICITY*.

Tests marked with * are not included within the scope of the ENAC accreditation



RESULTS

Carried out on the following reference:

MASK REF: PTXMASKV2_HOSP

With the performance requirements of EN 14683: 2019 + AC: 2019 standard for surgical masks points 5.2.2, 5.2.3 and 5.2.4 for types I, II and IIR.

Having obtained the following results:

| Operating requirements: | | | | | RESULTS (Average ± SD) |
|-------------------------|--|--------------|--------------|----------|----------------------------|
| | | Type I | Type II | Type IIR | |
| 5.2.2. | Bacterial Filtration Efficiency (BFE) (%) | ≥ 95 | ≥ 98 | ≥ 98 | 99,81 ± 0,19 |
| 5.2.3. | Breathability: Differential pressure (Pa/cm ²) | < 40 | < 40 | < 60 | 54 |
| 5.2.4. | Splash resistance pressure (kPa) | Not required | Not required | ≥ 16 | 0 de 32 at 21.3 kPa |

Notes:

- The rest of the standard tests not indicated in this test report have not been evaluated.
- SD: Standard Deviation.



RESULTS

IN VITRO DETERMINATION OF BACTERIAL FILTRATION EFFICIENCY (BFE)*

Standard

EN 14683:2019+AC:2019

Test date

18/03/2020 – 20/03/2020

Batch n^o(1)

Sample reference

MASK REF: PTXMASKV2_HOSP

Number of test specimen

5

Size of test specimen

10 cm x 10 cm

Tested area of the test specimen

50 cm²

Description of the test specimen

Inner side to the aerosol challenge

Test environmental conditions

T^a 20 °C Hr 30 %

Test control unit

Six stage Andersen Sampler

Flow of air

28.3 l/min

Test microorganism

Staphylococcus aureus ATCC 6538

_____>>>



RESULTS

Incubation conditions

48 h at 36 ± 1 °C

Test time

2 min / test specimen

Results

| Control values | | | | | | | |
|----------------|------------------------|------------------------|------------------------|------------------------|------------------------|------------------------|----------------------|
| | Level 1 (cfu/plate) | Level 2 (cfu/plate) | Level 3 (cfu/plate) | Level 4 (cfu/plate) | Level 5 (cfu/plate) | Level 6 (cfu/plate) | Total count (ufc) |
| C.P. | 66 | 64 | 272 | 196 | 144 | 29 | 771 |
| C.N. | 0 | 0 | 0 | 0 | 0 | 0 | 0 |

| Test sample values | | | | | | | |
|--------------------|------------------------|------------------------|------------------------|------------------------|------------------------|------------------------|----------------------|
| | Level 1 (cfu/plate) | Level 2 (cfu/plate) | Level 3 (cfu/plate) | Level 4 (cfu/plate) | Level 5 (cfu/plate) | Level 6 (cfu/plate) | Total count (ufc) |
| 1 | 0 | 0 | 0 | 0 | 0 | 0 | 0 |
| 2 | 1 | 0 | 0 | 0 | 0 | 0 | 1 |
| 3 | 1 | 0 | 0 | 0 | 1 | 1 | 3 |
| 4 | 0 | 0 | 0 | 0 | 0 | 0 | 0 |
| 5 | 1 | 0 | 0 | 1 | 0 | 1 | 3 |

Legend meaning:

cfu: colony forming units

C.P.: positive control (test run without test specimen).

C.N.: negative control (test run without bacterial suspension).

_____>>>



RESULTS

Calculation of bacterial filtration efficiency:

Equation: $B=(C-T)/Cx100$

:

C: Mean of the total plate counts for the two positive control runs

T: Total plate count for the test specimen

| Test | Filtration efficiency |
|-------------|-----------------------------------|
| 1 | 99,99 |
| 2 | 99,87 |
| 3 | 99,61 |
| 4 | 99,99 |
| 5 | 99,61 |
| Mean | 99,81 ± 0,19⁽²⁾ |

Notes

- The performance requirement for surgical mask according with EN 14683:2019+AC:2019, is:

| Test | Type I | Type II | Type IIR |
|---|--------|---------|----------|
| (BFE) % Bacterial filtration efficiency | ≥ 95 | ≥ 98 | ≥ 98 |

- Tested samples were supplied by the customer.

- ⁽¹⁾Data provided by the customer.

- ⁽²⁾ Standard Deviation of the results.

_____///



RESULTS

DETERMINATION OF BREATHABILITY (DIFFERENTIAL PRESSURE)*

Standard

EN 14683:2019+AC:2019

Principle

It is measure the differential pressure required to move air through a measured surface area at a constant flow of air, with the aim of measuring the pressure of air exchange of the material of the surgical mask.

Test date

20/03/2020 – 21/03/2020

Batch n^{o(1)}

Sample reference

MASK REF: PTXMASKV2_HOSP

Number of test specimen

5

Size of test specimen

4.9 cm²

Tested area of the test specimen

Circular, diameter 2.5 cm

Test environmental conditions

T^a 21 °C Hr 30 %

Flow of air

(8 ± 0.2) l/min

_____>>>



RESULTS

Results

| Test specimen | Pos1 Pa | Pos2 Pa | Pos3 Pa | Pos4 Pa | Pos5 Pa | Average Pa | ΔP (Pa/cm ²) |
|----------------|------------|------------|------------|------------|------------|---------------|-------------------------------------|
| 1 | 268 | 268 | 259 | 263 | 241 | 260 | 53 |
| 2 | 261 | 253 | 276 | 264 | 263 | 263 | 54 |
| 3 | 262 | 254 | 270 | 275 | 264 | 265 | 54 |
| 4 | 262 | 226 | 272 | 266 | 292 | 264 | 54 |
| 5 | 276 | 265 | 256 | 265 | 250 | 262 | 54 |
| Average | | | | | | 263 | 54 |

Notes

- The performance requirement for surgical mask according with EN 14683:2019+AC:2019, is:

| Test | Type I | Type II | Type IIR |
|--|--------|---------|----------|
| Differential pressure (Pa/cm ²) | < 40 | < 40 | < 60 |

- Tested samples were supplied by the customer.

⁽¹⁾Data provided by the customer.

///



RESULTS

DETERMINATION OF PRESSURE OF SPLASH RESISTANCE*

Standard EN 14683:2019+AC:2019 **Test method** ISO 22609:2004

Principle:

A defined volume of synthetic blood is shot with defined speeds of a pneumatically checked valve at the test specimen, in order to simulate a squirting of blood and other body fluids for the sample material.

The speeds and the selected volume correspond to a certain blood pressure, which spurts out by a defined opening size. The test could be performed with a pressure of 80, 120 and 160 mmHg. The back of the mask is examined by means of visual inspection and swab on penetrating liquid.

120 mmHg corresponds to the average systolic arterial blood pressure. The more the resistance against liquid splashes, the more barrier is the liquid resistance.

Test date

19/03/2020 – 20/03/2020

Batch n^{o(1)}

Sample reference

MASK REF: PTXMASKV2_HOSP

Number of test specimen

32

Size of test specimen

Circular diameter 5 cm

Tested area of the test specimen

19.6 cm²

Conditioning T^a 22 °C Hr 80 % **Test environmental conditions** T^a 20 °C Hr 32 %

Test parameters 21,3 kPa (160 mm de Hg) **Volume of synthetical blood** 2.0 mL

_____>>>



RESULTS

Results

| Pressure 21,3 kPa (160 mm de Hg) | | |
|----------------------------------|--------|------|
| Replica | Passed | Fail |
| 1 | X | |
| 2 | X | |
| 3 | X | |
| 4 | X | |
| 5 | X | |
| 6 | X | |
| 7 | X | |
| 8 | X | |
| 9 | X | |
| 10 | X | |
| 11 | X | |
| 12 | X | |
| 13 | X | |
| 14 | X | |
| 15 | X | |
| 16 | X | |
| 17 | X | |
| 18 | X | |
| 19 | X | |
| 20 | X | |
| 21 | X | |
| 22 | X | |
| 23 | X | |
| 24 | X | |
| 25 | X | |
| 26 | X | |
| 27 | X | |
| 28 | X | |
| 29 | X | |
| 30 | X | |
| 31 | X | |
| 32 | X | |

>>>



RESULTS

Remarks

- To pass the test no more than 3 samples at each pressure may fail.
- The performance requirement for surgical mask according with EN 14683:2019+AC:2019, is:

| Test | Type I | Type II | Type IIR |
|-------------------------------------|--------------|--------------|-----------|
| Pressure of splash resistance (kPa) | Not required | Not required | ≥ 16 |

///



RESULTS

DETERMINATION OF A POPULATION OF MICROORGANISMS ON PRODUCTS

Standard

EN 14683:2019+AC:2019; EN ISO 11737-1:2018

Reference

MASK REF: PTXMASKV2_HOSP

Batch number ⁽¹⁾

Sample size

3,29 g

Replica number

5

Test date

20/03/2020 – 27/03/2020

Test equipments

Incubator (03068E05) and Incubator (03202E05)

Results

| Parameter | Replica 1 (cfu/g) | Replica 2 (cfu/g) | Replica 3 (cfu/g) | Replica 4 (cfu/g) | Replica 5 (cfu/g) | Average (cfu/g) |
|-------------------------------|----------------------|----------------------|----------------------|----------------------|----------------------|--------------------|
| Aerobic bacteria to 33 ± 2°C | <1 | <1 | <1 | <1 | <1 | <1 |
| Moulds and yeasts to 22 ± 2°C | <1 | <1 | <1 | <1 | <1 | <1 |



RESULTS

Notes

⁽¹⁾Data provided from customer

The total count of microorganisms in the sample is <2 cfu/g

In accordance with the standard EN 14683:2019+AC:2019, the results must be in the values of the following table:

| Parameter | Units | Requirement |
|-----------------------|-------|-------------|
| Cleanliness microbial | cfu/g | ≤ 30 |



RESULTS

TEST FOR CYTOTOXICITY*

Standard

EN ISO 10993-5:2009

Test method

Direct contact

Exposure period

24 h.

Culture plates

EMEM

Celular line

NCTC-L-929

Test date

23/03/2020 – 25/03/2020

Reference

MASK REF: PTXMASKV2_HOSP

Quatitative evaluation**- Sample**

99 % Viable cells (Vital stain: Trypan Blue)

- Negative control

100 % Viable cells (Vital stain: Trypan Blue)

_____>>>



RESULTS

Qualitative evaluation

After the contact period are not observed structural alterations in the cell monolayer under or around the sample. The cells maintain cell membrane integrates and there is no evidence or other alteration cytoplasmic vacuolization suggestive of cell damage.

Conclusion **Grade 0**

TABLE 1. Evaluation of the cytotoxicity grade for the qualitative evaluation of the direct contact test.

| Cytotoxicity grade | Reactivity | Description of the reactivity zone |
|--------------------|----------------|---|
| 0 | Non reactivity | Zone non detectable around or under the sample. |
| 1 | Light | Some malformed or degenerated cells below the sample. |
| 2 | Slight | Zone limited to area under sample. |
| 3 | Moderate | Zone extending up to 1 cm from the edge of the sample size. |
| 4 | Severe | Zone extending more than 1 cm from the edge of the sample size. |

Remarks

- In the quantitative assessment, a value of less than 70% of viable cells was considered cytotoxic effect.
- In the qualitative assessment is considered like cytotoxic effect, grade higher to 2 in Table 1.



Judit Sisternes
Head of Health & Hygiene Products Division



Digitally signed by JUDIT SISTERNES
NAVARRO - NIF:48292160A
Date: 2020.03.31 15:00:32 +02:00
Reason: Autorizado
Location: Alcoy

LIABILITY CLAUSES

- 1.- AITEX is liable only for the results of the methods of analysis used, as expressed in the report and referring exclusively to the materials or samples indicated in the same which are in its possession, the professional and legal liability of the Centre being limited to these. Unless otherwise stated, the samples were freely chosen and sent by the applicant.
- 2.- AITEX shall not be liable in any case of misuse of the test materials nor for undue interpretation or use of this document
- 3.- The Offer and / or Order to which the applicant gives approval through signature and seal, constitutes the Legally Executable Agreement in which AITEX is responsible for safeguarding and guaranteeing the absolute confidentiality of the management of all the information obtained or created during the performance of the contracted activities.
- 4.- In the eventuality of discrepancies between reports, a check to settle the same will be carried out in the head offices of AITEX. Also, the applicants undertake to notify AITEX of any complaint received by them as a result of the report, exempting this Centre from all liability if such is not done, the periods of conservation of the samples being taken into account.
- 5.- AITEX is not responsible for the information provided by customers, which is reflected in the Report, and may affect the validity of the results.
- 6.- AITEX will provide at the request of the person concerned, the treatment of complaints procedure.
- 7.- AITEX is not responsible for an inadequate state of the sample received that could compromise the validity of the results, expressing such circumstance, in the test reports.
- 8.- AITEX may include in its reports, analyses, results, etc., any other evaluation which it considers necessary, even when it has not been specifically requested.
- 9.- When a Declaration of Conformity is requested, if not indicated otherwise, the decision rule will be applied according to ILAC-G8 & ISO 10576-1, in case of ambiguity, or indeterminacy
- 10.- The uncertainties of tests, which are made explicit in the Results Report, have been estimated for a $k = 2$ (95% probability of coverage). If not informed, they are available to the client in AITEX.
11. - The original materials and rests of samples, not subject to test, will be retained in AITEX during the twelve months following the issuance of the report, so that any check or claim which, in his case, wanted to make the applicant, should be exercised within the period indicated.
- 12.- This report may only be sent or delivered by hand to the applicant or to a person duly authorised by the same.
- 13.- The results of the tests and the statement of compliance with the specification in this report refer only to the test sample as it has been analyzed / tested and not the sample / item which has taken the test sample.
- 14.- The client must attend at all times, to the dates of the realization of the tests.
- 15.- According to Resolution EA (33) 31, the test reports must include the unique identification of the sample, and any brand or label of the manufacturer may be added. It is not allowed to re-issue test reports of untested sample names (references), they can only be re-issued for error correction or inclusion of omitted data that were already available at the time of the test. The laboratory can not assume responsibility for declaring that the product with the new trade name / trademark is strictly identical to the one originally tested; This responsibility belongs to the client.